

Memo

date: February 1, 2013

to: Thomas Roser

from: Yousef Makdisi (on behalf of the Experiment Safety Review Committee)

*subject: Recommendation for approval to introduce flammable gas into the PHENIX
Detector.*

The PHENIX detector was subjected to an annual ESRC review on December 20, 2012 and relevant action items are closed. The Walkthrough is scheduled to take place. The PASS flammable gas detectors and IR ventilation fans have been certified per an email from J. Reich. The Blue Sheeting was completed and signed off on January 30 2013. The crane is locked out and a shift schedule is in place. The shield wall is not complete but in discussion with Ray Karol we decided this is not a requirement.

With this, approval is requested to introduce flammable gases into the PHENIX detector systems.

All relevant documentation is in the C-AD Experiment Safety Review Committee files.

cc:

E. O'Brien
P. Ingrassia
Y. Makdisi
J. Nagel/D. Morrison
D. Phillips
P. Pile
B. Mueller

File

Approved: Thomas Roser

Date: 2/1/13